

# Standard Test Method for Indentation Hardness of Aluminum Alloys by Means of a Webster Hardness Gage<sup>1</sup>

This standard it issued under the fixed designation B647; the number immediately following the designation indicates the year of original adoption or, in the case of revision, the year of last revapion. A number in parentheses indicates the year of last reapproval. A superscript epision (e) indicates an editorial change since the last revision or reapproval.

### 1. Scope

- 1.1 This test method covers the determination of indeptation hardness of aluminum alloys with a Webster hardness gave. Model B.
- 1.2 The values stated in SI units are to he regarded as standard. No other units of measurement are included in this standard.
- 1.2.1 Exception—The values given in parentheses are for information only.
- Nore I—Two other models, A and B-75, are in use, but are participated in this test method. Model A does not provide numerical values of hardness and Model B-75 covers only a part of the range of numerical aluminum alloys.
- 1.3 This standard does not proper to entries off of the safety concerns, if any, associated with 132 Mer. It is the responsibility of the user of UN Sadndard to MADISH appropriate safety and health pradicts and determine the applicability of regulatory limitatives proper to use.

# 2. Referenced Documents

2.1 ASTM Standards:2

E6 Terminology Relating to Methods of McChanical Testing E10 Test Method for Brinell Hardness of Metolic Mackinsk E18 Test Methods for Rockwell Hardness of Metolic Materials

E177 Practice for Use of the Terms Precision and Bias in ASTM Test Methods

E691 Practice for Conducting an Interlaboratory Study to Determine the Precision of a Test Method

### 3. Terminology

3.1 Definitions—The definitions of terms relating to hardness testing appearing in Terminology E6 shall be considered as applying to the terms used in this test method.

## 4. Significance and Use

- 4.1 The Webster hardness gage is portable and therefore useful for in situ determination of the hardness of fabricated parts and individual test specimens for production control purposes. It is not as sensitive as Rockwell or Brinell hardness machiness ee 10.2.
- 42:516s test method should be used only as cited in applicable tracerial specifications.

# 5. Apparatus (Fig. 1)

- 5.1 The Webster hardness gage. Model B. consists of three morphirts: the frame, operating handle, and penetrator housing assembly includes the principal working parts, including the penetrator, loading spring, adjusting nut, penetrator housing begins key, return spring, and dial indeador.
  - 5.2 The indentor is a hardened steel truncated cone.
- 5.3 The dial indicator is graduated from 1 to 20, and is actuated by the penetrator so that the higher the reading, the higher is the hardness of the test material.
- 5.4 The configuration of the Webster hardness gage is such that it is operated like a pair of pliers.
- 5.5 The clearance between the penetrator and the anvil is about 6 mm ( $\frac{1}{4}$  in.), limiting the thickness of sample that can be tested.

## 6. Test Parts or Specimens

- 6.1 Any part or piece of material greater than 1 mm (0.04 in.) in thickness and equal to or less than 6 mm ( $^{1/4}$  in.) in thickness and with a clear flat area at an edge approximately 25 by 25 mm (1 by 1 in.) in size is suitable for test.
- 6.2 The surfaces shall be essentially parallel, smooth, clean, and free of mechanical damage. The test surface may be lightly polished to eliminate scratches or die lines.
- 6.3 The clear, flat area shall be such that there will be a clear distance of at least 3 mm (½ in.) from the edge of the part or specimen.

<sup>&</sup>lt;sup>1</sup> This test method is under the jurisdiction of ASTM Committee B07 on Light Metals and Alloys and is the direct responsibility of Subcommittee B07.05 on Testing.

Current edition approved May 1, 2016. Published May 2016. Originally approved in 1984. Last previous edition approved in 2010 as B647-10. DOI: 10.1520/B0647-10R16.

<sup>&</sup>lt;sup>2</sup> For referenced ASTM standards, visit the ASTM website, www.astm.org, or contact ASTM Customer Service at service@astm.org. For Annual Book of ASTM Standards volume information, refer to the standard's Document Summary page on the ASTM website.






FIG. 1. Webster Hardness Gage, Medel B

6.4 Parts or specimens with a slight tapes of subvalure may also be tested if a round anvil as used, as described in 8.1.1 and Fig. 2.

### 7. Calibration

another, or

- 7.1 Zero Adjustment:
- 7.1.1 Operate the instrument against the base anvil and note whether or not the indicator gives the zero (full-scale) restling.
  7.1.2 A correction in zero reading is nonpally not needed.
- except for one of the reasons listed below:
- 7.1.2.3 Excessive wear has taken place.
- 7.1.3 If an adjustment in zero is needed, turn the zero dijustment screw below the indicator dial slowly (Fig. 3), while operating the tester against the bare anvil and maintaining handle pressure, until the hand of the dial indicator rests on the zero line (which is located at the full-scale reading of 20).






FIG. 3 Zero Adjustment Screw, A (indicated by arrow)

- 7.1.4 The zero adjustment should never be used to make the indicator read correctly on a standard sample.
  - 7.2 Load Spring Adjustment:
- 7.2.1 Measure the hardness of a standard sample of medium to low hardness, either one furnished with the tester or one developed from reference stock; the dial should show the reading indicated for the standard within ±0.5.
- 7.2.2 If the readings on the standard or the bare anvil, or both, do not indicate the proper values, adjust the load spring with the special wrench provided until agreement is reached (Fig. 4).

## 8. Procedure

8.1 Place the test part or specimen between the penetrator and the anvil, and apply pressure to the handle.



FIG. 4 Adjusting Load on Load Spring

- 8.1.1 If the test part or specimen has a slight 15642 secureature, use a round anvil and apply 166 gage 1 secure 6.00 with the penetrator is applied normal tender (e.g. surface and the anvil bears along a line that is particle of 1747 specimen outside with the penetrator, as illustrated 1017 ig 25.
- 8.2 Apply sufficient pressure to cause the fine face of the penetrator housing to come in applicact with the surface of the test part or specimen.
- Note 2—Excess pressure on the handle is not harmful and does not affect the reading, but neither is it necessary, as the hardness reading is determined solely by the spring deflection.
- 8.3 Read the hardness from the dial indicator. Report the reading to the nearest 0.5.

Note 3—For relatively soft materials, the dial may indicate 30th drift toward lower numbers with time, after the initial pressure: 10 precommended that readings be made quickly and that the highest observed value be used.

### 9. Report

- 9.1 The report shall include the following:
- 9.1.1 Identification of material tested,
- 9.1.2 Model and serial number of hardness gage,
- 9.1.3 Number of readings taken,
- 9.1.4 Average of hardness values, rounded to the nearest half division, and
  - 9.1.5 Date of test.

### 10. Precision and Bias

- 10.1 The Webster hardness gage, Model B, is useful for hardness measurements of material in the range from 3003-0 to 7075-T6 representing a range in Rockwell hardness from about 5 HRE to 110 HRE.
- 10.2 The application of portable impressors, such as the Webster, will produce greater variation in hardness readings than standard fixed frame procedures, such as Test Methods E10 and E18. Further, since the scale of the Webster hardness age, Model B, has 20 divisions compared to 110 for the Rockwell E scale, it is also less sensitive than the Rockwell scale.
- 10.3 The variation in readings that results from the application of the Webster hardness impressor has not yet been established.
- 10.4 The precision of this test method is based on an 10.4 The precision of this test method is based on an laboratories tested a total of four alloys. Every "test result" precision the average of three individual determinations.
- Three-engings (9 total test readings) from every laboratory were analyzed in order to determine the precision statistics Metal behat; Practice E991 was followed for the design and analysts of the data; the details are given in ASTM Research Report No. 507-1001.
- 4051 Repeatability limit (r)—Two test results obtained retire one laboratory shall be judged not equivalent if they third by more than the "r' value for that material; "r' is the interval representing the critical difference between two test results for Die Same material, obtailed by the same operator using the same equipment on the same day in the same laboratory;
  - 10.4.1.1 Repeatability limits are listed in Table 1 below.
- 10.4.2. Reproduct hility limit 270—Two test results shall be judged not equivalent if they differ by more than the "R" value for that material: "R" is the interval representing the critical difference between two test results for the same material, obtained by different operators using different equipment in different laboratories.
  - 10.4.2.1 Reproducibility limits are listed in Table 1 below. 10.4.3 The above terms (repeatability limit and reproduc-
- ibility limit) are used as specified in Practice E177. 10.4.4 Any judgment in accordance with 10.4.1 and 10.4.2 would normally have an approximate 95 % probability of being correct, however the precision statistics obtained in this ILS

TABLE 1 Summary Statistics for Webster Hardness (Webster Scale Units) From Tests On Several Materials

Material	Average <sup>A</sup>	Repeatability	Reproducibility	Repeatability	Reproducibility	
		Standard Deviation	Standard Deviation	Limit	Limit	
	×	S,	S <sub>R</sub>	r	R	
3003-02	4.639	0.078	1.128	0.218	3.157	_
5052-H32	10.356	0.546	1.665	1.529	4.663	
6061-T6	16.172	0.286	0.380	0.801	1.063	
7075-T6	19.028	0.096	0.096	0.269	0.269	

AThe average of the laboratories' calculated averages.

establica de la contraction de

<sup>&</sup>lt;sup>3</sup> Supporting data have been filed at ASTM International Headquarters and may be obtained by requesting Research Report RR:B07-1001.

must not be treated as exact mathematical quantities which are applicable to all circumstances and uses. The limited number of materials tested and laboratories reporting results guarantees that there will be times when differences greater than predicted by the ILS results will arise, sometimes with considerably greater or smaller frequency than the 95 of probability limit would imply. Consider the repeatability limit and the reproducibility limit as general guides, and the associated probability of 95 % as only a rough indicator of what can be expected.

10.5 Bias—At the time of the study, there was no accepted reference material suitable for determining the bias for this test method, therefore no statement on bias is being made.

10.6 The precision statement was determined through statistical examination of 60 results, from five laboratories, on four alloys, described below:

3003-02 - Alcoa Rolled Plate 0.250" x 12" x 12" from 48" wide production 5552-H32- Alcoa Rolleri Plate 0.250" x 12" x 12" from 48"

wide production 6061-T6 - Kaiser Rolled Plate 0.250" x 12" x 12" from 48" wide production

7075-T6 - Kaiser Rolled Plate 0.250" x 12" x 12" from 48" wide production

ASTM International takes no position respecting Mo. baldily of any patent rights asserted in connection with any item mentioned in this standard. Users of this standard are expressly advised that determination of the validity of any such patent rights, and the risk of intringement of such rights, are entirely their dark explainability.

This standard is subject to revision at 40 IMA by the \$\frac{1}{2}

